

 Teco Diagnostics Clinical Reagent Specialists	Form	QL801-058
	Revision	A
	Prepared by	Brian Anderson
	Approved by	Owen Bry
	Issue date	1/14/2011
[MSDS] One-Step Urine Card Test		
1268 N. Lakeview Ave. Anaheim, CA 92807 Phone: (714) 463-1111 Fax: (714) 463-1169 www.tecodiag.com		

Section I - Name and Product

<i>Address (Street, State, and Zip Code)</i>		
1268 N. Lakeview Ave., Anaheim, California 92807 USA		
<i>Product Name</i>	<i>Part No.</i>	<i>Contact</i>
One-Step Urine Card Test	P594-25, P594-50	Quality Assurance
<i>Chemical Family or Product Type</i>	<i>Date of Preparation</i>	<i>Emergency Telephone No.</i>
Rapid Tests	Revised 01/14/2011	Weekdays 8:00-4:30 PST Tel: (800)222-9880 or (714) 463-1111 Fax#: (714) 463-1169

Section II - Composition

<i>CHEMICAL NAME</i>	<i>%</i>	<i>CAS#</i>
Buffered Salts with Antibody dye	Mixture	N/A
Other components either non-hazardous or at concentrations below that requiring hazardous listing.		

Section III - Physical Data

<i>Boiling Point</i>	<i>Melting Point</i>	<i>Specific Gravity</i>
N/A	N/A	N/A
<i>Vapor Pressure</i>	<i>Percent Volatile by Vol.</i>	<i>Vapor Density</i>
N/A	N/A	N/A
<i>Evaporation Rate</i>	<i>Solubility in Water</i>	<i>Appearance/Odor</i>
N/A	Insoluble	Cotton Pad

Section IV - Fire and Explosive Hazard Data

<i>Flash Point (Method Used)</i>	<i>Flammable Limits</i>	<i>LEL</i>	<i>UEL</i>
N/A	N/A		
<i>Extinguishing Media</i>			
Use media appropriate for surrounding materials.			
<i>Special Fire Fighting Procedures</i>			
N/A			
<i>Explosion Potential</i>			
The device contains sodium azide which may react with lead or copper plumbing to form potentially explosive metal azides.			

Section V - Health and Hazard Data

<i>Threshold Limit Value (TLV)</i>
N/A
<i>Effects of Overexposure</i>
None
<i>Emergency and First Aid Procedures</i>
N/A



Form	QL801-058
Revision	A
Prepared by	Brian Anderson
Approved by	Owen Bry
Issue date	1/14/2011

[MSDS] **One-Step Urine Card Test**
1268 N. Lakeview Ave. Anaheim, CA 92807 Phone: (714) 463-1111 Fax: (714) 463-1169 www.tecodiag.com

Section VI - Reactivity Data

Stability: Unstable Stable <input checked="" type="checkbox"/>	Polymerization: May Occur Will Not Occur <input checked="" type="checkbox"/>
<i>Incompatibility (Materials to avoid)</i> N/A	
<i>Decomposition Products</i> N/A	
<i>Conditions to be avoided</i> Avoid high temperature, high humidity conditions. Avoid prolonged exposure to air and light.	

Section VII - Spill and Disposal Procedures

<i>Steps to be taken in case of Leaks or Spills</i> Sweep up or pick up and place in a disposal container.
<i>Waste Disposal Method</i> Follow Federal, State and local regulations.

Section VIII - Personal Protection Information

<i>Respiratory Protection (Type)</i> Respiratory protection is not required.	
<i>Ventilation</i>	<i>Mechanical</i> Good General Ventilation.
	<i>Local Exhaust</i> N/A
<i>Eye Protection</i> Eye protection is not required to use this product	<i>Protective Gloves</i> Protective gloves recommended.
<i>Other Protective Equipment</i> Lab coat recommended.	
<i>Measures to be taken for equipment that has been in contact with the material</i> After ensuring equipment is disconnected from power supply, wipe the equipment with a clean lint-free cloth or towel moistened with water. Dispose the used cloth as in VII above.	

Section IX - Special Precautions

<i>Precautions to be taken in Handling and Storage</i> Store at temperature listed on product label.
<i>Other Precautions</i> Do not pipette by mouth. Avoid contact or ingestion.

Section X - Other Information

Proper Shipping Name: None; not regulated. Non-Hazardous for Transport: This substance is considered to be non-hazardous for transport. Non-Hazardous for Air Transport: Non-hazardous for air transport. This product is labeled in accordance with CFR21 (Food and Drugs), Section 809.10.

The information contained herein has been compiled from data presented in various technical sources believed to be accurate. Teco Diagnostics makes no warranties and assumes no liability in connection with the use of this information. It is the user's responsibility to determine the suitability of this information and to assure the adoption of necessary safety precautions.

N/A - Not Applicable or Not Available

TECO DIAGNOSTICS