



TECO DIAGNOSTICS

1268 N. Lakeview Ave.
Anaheim, CA 92807
1-800-222-9880

ALBUMIN – BCG METHOD TC MATRIX

INTENDED USE

For the quantitative determination of albumin in serum or plasma on TC Matrix analyzers.

SUMMARY AND EXPLANATION OF THE TEST

Albumin is the most abundant individual protein in serum. The reference range for serum albumin in an adult (>18-60 years) is 35-50 g/L (3.5-5.0 g/dL).⁶ The liver is the site of synthesis of albumin and hence liver disease is one of the most common causes of decreased albumin level. Other conditions such as congestive heart failure, malnutrition, starvation, nephritis, glomerulonephritis, leukemia and severe burns also resulted in the decrease of albumin level.

Albumin levels may be elevated in dehydration, shock, hemoconcentration, or upon administration of large quantities of concentrated albumin intravenously.

Bromocresol green (BCG), an ionic dye, binds tightly to albumin when serum is added. The albumin-BCG complex absorbs light at 600 nm more intensively than BCG alone.

The TC Matrix System automatically proportions the appropriate sample and reagent volumes into the cuvette. The system monitors the change in absorbance at 600 nanometers. This change in absorbance is directly proportional to the concentration of albumin in the sample and is used by the TC Matrix System to calculate and express albumin concentration.

Albumin + BCG -----> Albumin-BCG complex.

REAGENT CONTENTS

Each kit contains: Six Albumin reagent (6 x 40 ml)
Instruction Insert

REAGENT PREPARATION

No preparation is required.

REAGENT COMPOSITION

Bromocresol green: 0.28 mmol/L

Also non-reactive chemicals for optimal system performance.

REAGENT STORAGE AND STABILITY

Albumin reagent stored unopened at room temperature is stable until the expiration date shown on the bottle label. Once opened, Albumin reagent is stable for 30 days, or until the expiration date on the label.

DO NOT FREEZE.

SPECIMEN COLLECTION AND HANDLING

1. The test can be performed on serum or plasma. For serum, blood is drawn into a tube which does not contain anticoagulant and is allowed to clot. The serum is then separated from the clot. A maximum limit of two hours from the time of collection is recommended.
2. Separated serum or plasma should not remain at room temperature longer than 8 hours. If assays are not completed within 8 hours, serum and plasma should be stored at 2°C to 8°C. If assays are not completed within 48 hours, or the separated sample is to be stored beyond 48 hours, samples should be frozen at -15°C to -20°C. Frozen samples should be thawed only once. Analyte deterioration may occur in samples that are repeatedly frozen and thawed.
3. For plasma, add whole blood directly into a tube containing anticoagulant. Acceptable anticoagulants are listed in the "LIMITATIONS" section.

CALIBRATION

1. Calibrator required: TECO MULTI Calibrator is needed but is not provided in the kit.
2. The system must have a valid calibration in memory before controls or patient samples can be run.
3. The TC Matrix system will automatically perform checks on the calibration and produce data at the end of calibration.

MATERIALS NEEDED BUT NOT SUPPLIED WITH REAGENT KIT

TECO MULTI Calibrator.

At least two levels of control material.

LIMITATIONS

1. The anticoagulants Potassium Oxalate, Sodium Fluoride and Sodium Citrate were found to be incompatible with this method.
2. The anticoagulants EDTA, Lithium Heparin, Sodium Heparin, Ammonium Heparin were found to be compatible with this method.

INTERFERENCE

1. Hemoglobin may interfere with this methodology.
2. Lipemic samples >3+ should be ultra-centrifuged and the analysis performed on the infranate.
3. On this method, refer to the work of Young for a review of drug and comprehensive list of substances which affect the Albumin level.¹⁰

EXPECTED VALUE

3.5-5.0 g/dL or 35-50 g/L

PRECAUTIONS

1. For in vitro diagnostic use only.
2. Since all specimens are potentially infectious, they should be handled with appropriate precautions and practices in accordance with Biosafety level 2 as recommended by USA NIH manual Biosafety in Microbiological and Biomedical Laboratories, and in accordance with National or local regulations related to the safety precautions of such materials.
3. Each laboratory should perform quality control testing to assure the results are reliable before testing the specimens.

PROCEDURES

TEST NAME:	ALB	R1:	450
TEST NO.		R2:	0
FULL NAME:	Albumin	SAMPLE VOLUME:	3
REFERENCE NO.:		R1 BLANK:	/
ANALY. TYPE:	Endpoint	MIX REAG. BLANK:	0.1 - 1.5
PRI. WAVE :	630 nm	CONCENTRATION:	/
SECON. WAVE:	/	LINEARITY LIMIT:	1 - 7
TREND:	Increase	SUBSTRATE LIMIT:	/
REACT. TIME:	0 - 18	FACTOR:	/
INCUBATE TIME:	/	PROZONE CHECK:	/
UNIT:	g/dl	Q1: / Q2: / Q3: / Q4: /	
PRECISION:	0.1	PC: / ABS.: /	
Calibration Type: Calibrate + Reag. Blank Calibration Rule: Two-point linear			

PERFORMANCE CHARACTERISTICS

Analytical Range: 1.0-7.0 g/dl

For albumin analysis using Teco Diagnostics' Albumin Reagent on TC Matrix System, this method has been demonstrated to be linear from 1.0-7.0 g/dl.

Accuracy: Comparison study was performed on TC Matrix System from 40 samples. Beckman Coulter Albumin reagent was used to compare with Albumin Reagent. The results of this study yield a correlation coefficient of 0.991 with a regression equation of $y=0.991X + 0.200$.

Precision: Within Run precision for Albumin Reagent was determined following a modification of NCCLS EP5-A. Two commercial human serum samples were assayed on TC Matrix System for 25 times.

Sample	Sample 1	Sample 2
N	25	25
Mean (g/dl)	4.0	2.8
Standard Deviation (g/dl)	0.09	0.07
Coefficient of Variation (%)	2.1	2.0

Run-Day precision for Albumin Reagent was determined following a modification of NCCLS EP5-A. Two commercial human serum samples were assayed on TC Matrix Systems five times per day for five days for the total of 25 values.

Sample	Sample 1	Sample 2
N	25	25
Mean (g/dl)	4.0	2.8
Standard Deviation (g/dl)	0.08	0.08
Coefficient of Variation (%)	2.1	2.2

REFERENCES

1. Doumas, B.T.et.al. Clin.Chim.Acta 31:87-96, 1971
2. Doumas, B.T.and Biggs, H.G. Standard Methods of Clinical Chemistry Vol. 7,175,1972.
3. Tietz, N.W., " Specimen Collection and Processing; Sources of Biological Variation," Textbook of Clinical Chemistry, 2nd Edition, W.B. Saunders, Philadelphia, PA (1994)
4. National Committee for Clinical Laboratory Standards. Approved Guideline. NCCLS publication C28-A, Villanova, PA (1994).
5. Henry, J. B., ed., Clinical Diagnostics and Management by Laboratory Methods, 18th Edition, W.B. Saunders, Philadelphia.
6. Tietz, N.W., ed., Clinical Guide to Laboratory Tests, 2nd Edition, W.B. Saunders, Philadelphia, PA (1990)
7. National Committee for Clinical Laboratory Standards. Method Comparison and Bias Estimation Using Patient Samples; Tentative Guideline, NCCLS Publication EP9-T, Villanova, PA (1993)
8. National Committee for Clinical Laboratory Standards, Precision Performance of Clinical Chemistry Devices; Tentative Guideline, 2nd Edition, NCCLS publication EP5-T2, Villanova, PA (1992)
9. National Committee for Clinical Laboratory Standards, National Evaluation Protocols for Interference Testing, Evaluation Protocol Number 7, Vol. 4, No. June 1984.
10. Young, D.S., Effects of Drugs on Clinical Laboratory Tests, 3rd. Ed., AACC Press, Washington DC, 1990, 3-104 thru 3-106.

A502-510TM: 03/2017

Manufactured by:



TECO DIAGNOSTICS
1268 N. LAKEVIEW AVE.
ANAHEIM, CA 92807
U.S.A.