

TECO DIAGNOSTICS

1268 N. Lakeview Ave.
Anaheim, CA 92807
1-800-222-9880

γ -GT LIQUID REAGENT (KINETIC METHOD)

INTENDED USE


For the Kinetic Quantitative Determination of γ -Glutamyl Transferase in Serum for Manual and/or Automated Procedures

SUMMARY AND PRINCIPLE

Gamma-glutamyl transferase (γ -GT) is one of a large group of enzymes known as peptidases. Although renal tissue has the highest level of γ -GT, the major source of the enzyme present in serum is of hepatic origin. Elevated levels of γ -GT are found in association with hepatobiliary and pancreatic disorders; alcoholics and heavy drinkers, in myocardial disorders and in diabetics¹.

Unlike alkaline phosphatase activity, the serum γ -GT activity remains normal in diseases affecting bone and during normal bone growth. Therefore, a rise in serum γ -GT activity may be considered as a sensitive and more specific indicator of liver disease than alkaline phosphatase activity.

The Teco γ -GT procedure has been optimized according to Szasz.



γ -GT catalyzes the transfer of a γ -Glutamyl group from L- γ -glutamyl-3-carboxy-4-nitroanilide. The rate of liberation of 5-amino-2-nitrobenzoate is directly related to the γ -GT activity in the sample and is quantitated by measuring the increase in absorbance at 405 nm.

REAGENTS

γ -GT Buffer (R1):

Tris, pH 8.25	100 mmol/L
Glycylglycine	100 mmol/L

γ -GT Substrate (R2):

L- γ -Glutamyl-3-Carboxy-4-Nitroanilide	4.0 mmol/L
--	------------

PRECAUTIONS

The reagents are for "In Vitro Diagnostic Use". Normal precautions exercised in handling laboratory reagents should be followed. The reagents contain sodium azide that may be toxic if ingested. Sodium azide may also react with lead and copper plumbing to form highly explosive metal azides. Refer to Material Safety Data Sheet for any updated risk, hazard or safety information.

REAGENT PREPARATION

Buffer and Substrate liquid reagents are supplied ready-to-use. Prepare Working Reagent in the ratio of 5 parts Buffer (R1) to 1 part Substrate (R2), (i.e., 25 mL Buffer and 5 mL Substrate).

REAGENT STORAGE AND STABILITY

Reagents are stable until the expiration date on their respective labels, when properly stored at 2-8°C and protected from light. R1 should appear clear/colorless while R2 should appear clear/yellow. Discard if either appears cloudy or contains particulate matter. The Working Reagent is stable for 4 weeks at 2-8°C or 5 days at room temperature

(15-25°C). The Working Reagent should be discarded if the initial absorbance, read against distilled water at 405 nm, is above 0.800.

MATERIAL REQUIRED BUT NOT PROVIDED

1. Spectrophotometer capable of absorbance reading at 405 nm and 1 cm light path
2. Constant temperature block/bath, 37°C, or temperature controlled cuvette
3. Accurate pipetting devices
4. Test tubes
5. Interval timer

SPECIMEN COLLECTION AND STORAGE

Serum or EDTA plasma, free of hemolysis, should be used. Complexing anticoagulants such as citrate, oxalate, fluoride and must be avoided since they inhibit γ -GT activity². The loss of γ -GT activity is minimized by storing the samples refrigerated for up to 7 days or frozen up to 2 months³. Bilirubin levels up to 40 mg/dL and triglyceride levels up to 2000 mg/dL show no interference in this test.

INTERFERING SUBSTANCES

γ -GT is an inducible enzyme. Consequently patients who are receiving antiepileptic drugs or aminopyrine show elevated γ -GT activity. Chronic use of ethanol also increases serum γ -GT activity⁴⁻⁵. Certain drugs and other substances are also known to affect γ -GT values⁶.

MANUAL PROCEDURE

1. Prepare γ -GT Working Reagent according to instructions.
2. Zero spectrophotometer at 405 nm with distilled water.
3. For each sample and control, add 1.0 mL Working Reagent to cuvette or test tube and warm to 37°C for 3 minutes.
4. Add 100 μ L (0.10 mL) serum to its respective tube and mix gently.
5. Read and record absorbance at 1 minute. Continue incubating at 37°C and record absorbance again at 2 and 3 minutes. Rate should be constant.
6. Determine the average absorbance per minute ($\Delta A/\text{min}$), multiply by factor 1158 for results in U/L.

NOTE: If cuvette is not temperature controlled, incubate samples at 37°C between readings.

AUTOMATED PROCEDURE

Special adaptations for automated analyzers are available by contacting Teco's Customer Service Department.

QUALITY CONTROL

It is recommended that controls be included in each set of assays. Commercially available control material with established γ -GT values may be used for quality control. The assigned value of the control material must be confirmed by this methodology. Failure to obtain the proper range of values in the assay of control material may indicate reagent deterioration, instrument malfunction or procedural errors.

CALIBRATION

γ -GT activity is based on the "micromolar extinction coefficient" of 5-amino-2-nitrobenzoate at 405 nm (see "Results" section). The instrument manufacturer's calibration guidelines should be followed to calibrate your analyzer.

RESULTS

Values are derived based on the "absorptivity micromolar extinction coefficient" of 5-amino-2-nitrobenzoate at 405 nm (0.0095). Units per liter (U/L) γ -GT activity are the amount of enzyme that produces one mmol/L of 5-amino-2-nitrobenzoate per minute.

$$U/L = \frac{\Delta A/Min}{Absorptivity} \times \frac{Total Volume}{Sample Volume}$$

$$U/L = \frac{\Delta A/Min}{0.0095} \times \frac{1.10}{0.10}$$

$$U/L = \Delta A/Min \times 1158$$

LIMITATIONS

If the $\Delta A/min.$ is greater than 0.259, dilute 1 part sample with 5 parts isotonic saline and re-assay. Multiply results by 6.

EXPECTED VALUES⁷

Normal Range: Males: 0 - 50 U/L (37°C)
Females: 0 - 30 U/L (37°C)

This range should serve only as a guideline. It is recommended that each laboratory establish its own range of expected values, since differences exist between instruments, laboratories, and local populations.

PERFORMANCE CHARACTERISTICS

Comparison: A group of 63 sera ranging in γ -GT activity from 15 - 714 U/L was assayed by the described γ -GT method and by a similar commercially available γ -GT reagent. Comparison of the results yielded a correlation coefficient of 1.000 and the regression equation was $y = 0.94x + 1.9$. (Comparison studies were performed according to NCCLS Tentative Guideline, EP9-T.)

Precision: Within-run precision was established by 20 assays on three different levels of commercial serum controls. Total Precision values were obtained by assaying the 3 commercial controls for 5 consecutive days.

	Within-Run		
	Serum 1	Serum 2	Serum 3
Mean γ -GT (U/L)	40	74	206
Std. Deviation (U/L)	1.0	0.9	1.3
C.V. (%)	2.5	1.2	0.6

	Total Precision		
	Serum 1	Serum 2	Serum 3
Mean γ -GT (U/L)	42	72	204
Std. Deviation (U/L)	0.6	0.6	0.7
C.V. (%)	1.5	0.9	0.4

Precision studies were performed according to NCCLS Tentative Guideline, EP5-T.

Linearity: Linear to 300 U/L at 37°C. Performed according to NCCLS Guideline EP6-P.

Sensitivity: Based on an instrument resolution of $A = 0.001$, the method presented shows a sensitivity of 1.0 U/L.

REFERENCES

1. Tietz, N.W. Fundamentals of Clinical Chemistry, 2nd ed. W.B. Saunders Co., Philadelphia, PA, p. 622 (1976)
2. Whitfield, JB, Moss DW, Neale, G, Orme, M and Breckenridge, A, Brit Med J 1, 316, 1973
3. Rosalki SB, Rau D, Lehman D, Prentice M: Determination of γ -glutamyltranspeptidase activity and its clinical applications. Ann Clin Biochem 7: 143, 1970
4. Rosalki SB, Tarlow D, Rau D: Plasma γ -glutamyltranspeptidase elevation in patients receiving enzyme inducing drugs, Lancet 2: 376 1971
5. Bartels H, Hauck W, Vogel I: aminopyrine an effective modifier of liver and serum γ -glutamyltranspeptidase. J Pediatr 86: 298, 1975
6. Young DS, Effects of drugs on clinical laboratory tests. AACC Press, Washington D.C. 1990
7. Tietz Textbook of Clinical Chemistry, 2nd ed. W.B. Saunders Co., Philadelphia, PA, p. 851 (1994)

G572: 08/2018

Manufactured by:

